

Series 228BR Metallic Tee Type Flow Sensors

DESCRIPTION

The Series 228 flow sensors from Badger Meter® feature a six-bladed impeller design with a proprietary non-magnetic sensing mechanism. The forward swept impeller shape provides higher, more consistent torque than four-bladed impeller designs and is less prone to be fouled by water-borne debris. The forward curved shape coupled with the absence of magnetic drag provides improved operation and repeatability, even at lower flow rates. This is especially true where the impeller is exposed to metallic or rust particles found in steel or iron pipes. As the liquid flow turns the impeller, a low impedance square wave signal is transmitted with a frequency proportional to the flow rate. The signal can travel up to 2000 feet between the flow sensor and the display unit without the need for amplification.

All sensors, except irrigation versions, are supplied with 20 feet of 2-conductor 20 AWG shielded UL type PTLC 221° F (105° C) cable.



The 228BR tee-mounted flow sensor consists of a standard 220BR sensor mounted in a 2 inch or 2.5 inch bronze tee.



SPECIFICATIONS

Wetted Materials (except tees)	See "Part Number Construction" on page 4				
Sensor Sleeve and Hex Adapter	Sleeve	Admiralty brass, UNS C44300			
Sensor Sieeve and nex Adapter	Hex Adapter	Lead-free brass, C89833			
Tee for 228BR	Cast bronze, Class 125 per ASME B16.15				
Towns and the Detinate	Standard Version	221° F (105° C) continuous service			
Temperature Ratings	Irrigation Version	150° F (66° C) continuous service			
	Temperature (F)	Pressure (psi)			
	-20150	200			
Pressure Ratings	200	190			
	250	180			
Recommended Design Flow Range	0.530 ft/sec				
Accuracy	±1.0% of full scale over recommended design flow range				
Repeatability	±0.3% of full scale over recommended design flow range				
Linearity	±0.2% of full scale over recommended design flow range				
	Supply voltage = 8V DC min. 35V DC max.				
	Quiescent current = 600 uA (typical)				
Transducer Excitation	OFF State (V_{High}) = Supply voltage – (600 $\mu \times$ Supply impedance)				
	ON State (V_{Low}) = 1.2V DC @ 40 mA (15 Ω + 0.7V DC)				
Electrical Cable for Standard Sensor Electronics	20 ft (6 m) of 2-conductor 20 AWG shielded UL type PTLC wire provided for connection to display or analog transmitter unit. Rated to 221° F (105° C). May be extended to a maximum of 2000 ft (610 m) with similar cable and insulation appropriate for application.				
Electrical Cable for IR Sensor Electronics	48 in. (122 cm) of UL Style 116666 copper solid AWG 18 wire with direct burial insulation. Rated to 221° F (105° C).				



DIMENSIONS

Standard Sensor




Figure 1: Standard 228BR flow sensor

	Series No. Complete	Tee No.	NPT Threads per inch	Α	B*	C*	D	E
228BR	228BR-2	71879T	211.5	4.25 in. (107.95 mm)	8.21 in. (208.53 mm)	6.75 in. (171.45 mm)	2.92 in. (74.17 mm)	6 in. (152.40 mm)
	228BR-2.5	71883T	2.58	4.81 in. (122.17 mm)	9.00 in. (228.60 mm)	7.25 in. (184.15 mm)	3.49 in. (88.65 mm)	6 in. (152.40 mm)

^{*} Dimensions (B, C) may vary ± 0.25 in., depending on the makeup of the pipe threads.

High Temperature Sensor





Figure 2: High temperature 228BR flow sensor

	Series No. Complete	Tee No.	NPT Threads per inch	A	B*	C *	D	E
228BR	228BR-2	71879T	211.5	4.25 in. (107.95 mm)	17.96 in. (456.18 mm)	16.50 in. (419.10 mm)	2.92 in. (74.17 mm)	6 in. (152.40 mm)
	228BR-2.5	71883T	2.58	4.81 in. (122.17 mm)	18.12 in. (460.25 mm)	16.37 in. (415.80 mm)	3.49 in. (88.65 mm)	6 in. (152.40 mm)


^{*} Dimensions (B, C) may vary ± 0.25 in., depending on the makeup of the pipe threads.

PART NUMBER CONSTRUCTION

Standard Sensor



High Temperature Sensor



Control. Manage. Optimize.

Data Industrial is a registered trademark of Badger Meter, Inc. Other trademarks appearing in this document are the property of their respective entities. Due to continuous research, product improvements and enhancements, Badger Meter reserves the right to change product or system specifications without notice, except to the extent an outstanding contractual obligation exists. © 2020 Badger Meter, Inc. All rights reserved.

www.badgermeter.com